

# Yorkshire and Humber Neonatal ODN Setting up a Patient on a Replogle Tube

Author Fiona Metcalfe

Date Written Jan 2019

Review Date Jan 2021

## Equipment

- Replogle tube –
  - Size 10Fg – greater than 1500g,
  - Size 8Fg - less than 1500g, for the very preterm infant
- Suction tubing X 2
- 1ml or 2ml Syringe
- Sodium chloride 0.9% for injection
- Hydrocolloid dressing (skin protection) Duoderm and fixation tape - Zinc oxide tape / Elastoplast
- Monitoring equipment for saturations and heart rate.
- **You need two Suction apparatus' available**
- One of which is a dedicated **low pressure** suction pump (**35-60cmH20/ 3.5-6kPa**)
- Resuscitation equipment, including 2<sup>nd</sup> suction apparatus for nasopharyngeal suctioning, suction catheters (various sizes), stethoscope, face mask, bagging and intubation equipment.
- Hand Mittens

## Procedure

Action	Rationale
<p>If oesophageal atresia (OA) suspected, apply necessary ICU monitoring</p> <p>And Nurse in Intensive care on full monitoring.</p>	<p>Infant with OA is at high risk of aspiration</p> <p>Patient needs constant monitoring to</p>

<b>One to one nursing care at all times</b>	prevent aspiration/choking & attend to Replogle tube
Assemble all equipment.	Need equipment available in case of sudden deterioration
Ensure <b>Two Suction apparatus</b> set up	One for <u>low pressure</u> suction of the replogle tube and one for, <u>normal suction pressure</u> , oral-pharyngeal suction  The low pressure MUST be labelled clearly
Clear oral and Naso-pharyngeal Secretions with suction and nurse head tilted up 30-45°.	To minimise risk of aspiration through the trachea-oesophageal fistula
Apply lubricant or sterile water to end of replogle tube	To assist passing the replogle tube
Gently advance Replogle tube, through the nose, until you feel resistance, then pull back 0.25cms. <b>(average length passed 10-12cms)</b>	Do this gently so as not to cause trauma or perforation and ensure optimal placement for clearing secretions  <b>NOTE: Give oro-pharyngeal suction while waiting for optimum position of replogle tube</b>
Explain what you are doing to parents and the rationale for this	To ensure parents are kept up to date
Connect Replogle tube to <b>low pressure Suction- 3.5-6kPa (35-60 cmsH2O)</b> <b>Can have Brief increase to 7-10kPa /</b> 70-100 cms H2O to clear thick secretions. Clearly label (Low Pressure Suction)	To allow continuous replogle Suction and not cause trauma.  To ensure correct suction used

Ensure secretions are draining <b>CONTINUOUSLY</b> along the replegle tube	<u>Continuous drainage demonstrates optimum tube placement</u>
Secure tube well using Duoderm on skin and then Zink oxide/Elastoplast tape -Put Mittens on hands if available	To protect skin and reduce risk of tube getting dislodged.
Document what the length of the tube inserted is at nose level	This will help if needing to replace replegle tube quickly.
<b>Flush 'side arm' 0.25-0.5ml Sodium Chloride 0.9%- every 10-15 mins (more often if needed)</b> <b>Flush 'large arm' every 30 mins.</b> <b>Do NOT leave syringe attached after instilling Sodium Chloride 0.9%</b>	To keep tube patent This will affect the drainage from the replegle tube <b>Flush <u>as often as needed</u> to clear tube.</b> <b>Can follow flush with air to clear tube</b>
<b>Following first insertion, check position.</b> Ensure secretions are draining constantly. For repeat insertions – check documented external length and insert to set position. Xrays are only indicated with problematic positions.	To ascertain position of Replegle tube/blind oesophageal pouch.
If you can't get secretions draining consider changing the tube.	Have spare tube available at bed side in case of blockage and emergency equipment
Note- Can put sign on bedside to highlight use of replegle tube, Size of tube and length inserted	To make it easier and quicker for staff to respond quickly to a tube blocking or getting dislodged.
Clearly mark the low pressure suction 'LOW' could also add the suction pressure range to a sticker on low pressure suction	To reduce the risk of using high pressure suction inappropriately.

Figure 1 Replogle tube




Figure 2  
Correct placement of Replogle tube




Diagram 4. The Replogle tube

Figure 3 Example of Low suction unit




Figure 4 Attachment of the Replogle tube to the low suction unit



Figure 5 Attachment of flush syringe to side port



DRAFT